



All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

Reference No : J-11058(001)/2016-17/S&P /4154

Dated : 17th Jan 2017

Corrigendum

Tender Reference : J-11058(001)/2016-17/S&P.
Last Date : 18-01-2017
Name of the item : Adult Cystoscope, Cystolitholapaxy Set, Visual Internal Urethrotomy (VIU) Set, Bipolar/Saline TURP Set and Bipolar Electro Cautery Set.
Last Date of Submission : 31st Jan 2017, 12.00 NOON
Tender opening date (Technical bid) : 31st Jan 2017, 03.00 PM

Revised Specification:

Read as Specification:

1. Forward oblique 12 degree telescope, 4mm diameter, autoclavable, length 30 cms, fibre optic light transmission incorporated – 1pc
2. Forward oblique 30 degree telescope, 4mm diameter, autoclavable, length 30 cms, fibre optic light transmission incorporated – 2pcs
3. Cystoscope-Urethroscope sheath, 19/19.5 Fr, with obturator and 2 Luer-lock connectors.
4. Cystoscope-Urethroscope sheath, 17/17.5 Fr, with obturator and 2 Luer-lock connectors.
5. Cystoscope-Urethroscope sheath, 22/23 Fr, with obturator and 2 Luer-lock connectors.
6. Telescope bridge with 1 lockable channel.
7. Optical Biopsy Forceps 7 Fr with double action jaws, flexible, length 40 cms – 1pc
8. Sachse Urethrotome Set – The set should have 21 Fr Urethrotome sheath with obturator and channel for Urethrotome working element. There should be a telescope bridge for use with the Urethrotome sheath with a 5 Fr channel for instruments. A supplementary half-moon sheath, that slips on the Urethrotome sheath and is side-open for introduction of the balloon-tipped

being provided with the resection set and should have cold knife, straight – 6nos and cold knife, half-round – 6 nos.

9. Maurmeyer Stone Punch: This set should be compatible with 4mm 12-degree telescope. The set should have the stone punch-working element and should be provided with compatible punch sheath with obturator. Sheath should have matt finish for staying of lubrication gel to reduce friction trauma to the patient.
10. Toomey glass syringe – 50 mL and 100 mL - 1 each in number.
11. Otis/Non-optical Urethrotome knife – should be supplied with the dilatation unit. Should be supplied with compatible knife and head parts.
12. Saline TUR Set

This system should be compatible with following procedures:

Saline/Bipolar Transurethral Resection

Conventional Monopolar Transurethral Resection

Bipolar/Saline Enucleation of Prostate

Open Monopolar cauterization

Open Bipolarcauterization

Laparoscopic - Monopolar

Laparoscopic – Bipolar

Should have inbuilt vessel sealing capability or should be upgradable to have open and laparoscopic vessel sealing.

System should have following features:

Output mode: Monopolar, Bipolar/Saline

Monopolar cutting: PURE, BLEND

Monopolar Coagulation: SPRAY

Bipolar Cutting: PURE

Bipolar Coagulation: SOFT, HARD

Saline cutting: PURE, BLEND

System should have inbuilt vessel sealing capability or should be upgradable to have vessel sealing in the future.

Saline coagulation: 2 or more different strengths

Base Frequency: 350 kHz

Protection against electric shock: Class I Type CF

Weight should not be more than 15 Kg.

Automatic Smoke Evacuation facility

System should include the following:

- Electrosurgical Unit – 1pc
- Footswitch – 1pc
- Bipolar Active Working Element – 1pc
- Monopolar Active Working Element – 1pc
- Bipolar HF-Cable – 1pc
- Monopolar HF-Cable – 1pc
- Rotatable Outer Sheath, 26 Fr./24Fr, 1/2 stopcock – 1 pc
- Resection.Inner Sheath, 24Fr/22 Fr with obturator and with/without stopcock – 1 pc ,
Inner Sheath Should be compatible with intermittent TUR
- Compatible visual Obturator-1

Should be supplied with compatible accessories/electrodes-

- Monopolar Cutting loop electrode (6 Pcs)
- Monopolar Coagulation roller electrode (6 Pcs)
- Bipolar Cutting Electrode (6 Pcs)
- Bipolar Needle Electrode (6 Pcs)
- Bipolar Button Electrode for plasma vaporization (2 Pcs)
- Bipolar Enucleation Electrode (6 Pcs)
- Light Guide Cable, 3.5 mm, 3 Meter

All instruments should be European CE or FDA certified wherever applicable.



**Administrative Officer
AIIMS, Bhubaneswar**

Copy to:

- | | | |
|------------------------|---|--------------------------------------|
| 1. PS to Director | : | for kind information to the Director |
| 2. PS to Fin. Advisor | : | for kind information of FA. |
| 3. PS to DD (Admin) | : | for kind information of DD (Admin). |
| 4. Med. Superintendent | : | for kind information Please. |
| 5. Dr.D.Bandhopahyaya | : | for hoisting in the website please. |