



All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

Reference No : J-11050(007)/2016-17/S&P

Dated : 25th Feb 2017

Corrigendum

Tender Reference : J-11050(007)/2016-17/S&P.

Last Date : 02-03-2017

Name of the item : **Supply & Installation of Cranial Trauma Neurosurgery Set.**

Last Date of Submission : 16th Mar 2017, 12.00 NOON

Tender opening date (Technical bid) : 16th Mar 2017, 03.00 PM

Addendum Specification (Chapter-III): In addition to existing Specification

Specification:

1. All the instruments should be made of high grade stainless steel / titanium.
2. The quality of steel should comply with international medical standards.
3. It should be of high quality and precision.
4. It should be light weight, strong and durable.
5. It should be non-magnetic even on repeated use.
6. All instruments should be autoclavable.
7. The Tungsten Carbide (TC) inlay should be nickel based.
8. The TC inlay should be welded to the instrument and not pasted.
9. All instruments should have bar coding printed on instruments.
10. All instruments should be supplied from single manufacturer.
11. The manufacturer should have a direct repair facility available in India.
12. All instruments should be autoclavable at 130 to 150 0 Celsius.
13. Instruments should meet the International Certifications such as DIN Or EN ISO 13485 Or USA FDA.
14. The Company should have Good Manufacturing Practices certificate of manufacturer with details of manufacturing facility mentioned.
15. Each instrument should be separately certified and should have sales permit in the country of manufacturing.

Asst. Store Officer
AIIMS, Bhubaneswar