



All India Institute of Medical Sciences, Bhubaneswar
At - Sijua (Patrapada), Post - Dumuduma, Bhubaneswar (Odisha) - 751 019

File No- STORE-113/17/2024-STPUR SEC/02

Dated: 20-05-2025

Corrigendum(e-Tender)-I

e-Tender Ref. No. : STORE-113/17/2024-STPUR SEC/02

Dated:15-04-2025

Name of the Tender : Supply and Installation &Commissioning of MRI-Compatible Syringe Pump at AIIMS, Bhubaneswar.

S.No.	S.No of NIT Technical Specifications	Existing technical Specification	To be read as
1	9	The pump should be capable of functioning standalone and being fixed on a platform/stand.	The workstation should be capable of functioning standalone with facility to stack pumps.
2	12	It should have a standardized interface with keypads	It should have a standardized interface.
3	13	The flow rate should be adjustable from 0.01 ml/hr to 999 ml/hr.	The flow rate should be adjustable from 0.1 ml/hr to 999 ml/hr
4	14	The flow rate can be adjustable in 0.01 ml/hr increments	The flow rate can be adjustable in 0.1 ml/hr increments.
5	21	The bolus rate can be adjusted from 1 to 1500 ml/ hr in steps of 0.01 ml	The bolus rate can be adjusted from 1 to 1500 ml/ hr in steps of 0.1 ml.
6	23	The bolus can be delivered without selecting inputs (by pressing a button).	The bolus can be delivered without selecting inputs (by pressing a button/icon on the screen).
7	27	Internal battery life should be at least 10 hours after full charge (@ 25 ml/ hr).	Internal battery life should be at least 10 hours after full charge (@ 5 ml/ hr).
8	34	The pump should be waterproof so that fluid does not enter the pump in case of accidental spillage (IP34 certified).	The pump should be waterproof so that fluid does not enter the pump in case of accidental spillage (IP33 or better certified).
9	Last date and time of submission of e-Tender	Dt.30-05-2025 at 05:00PM	Dt.04-06-2025 at 05:00Pm
10	Date & time of opening of e-Tender	Dt.05-06-2025 at 03:00PM	Dt.10-06-2025 at 03:00PM

Note-All other terms and conditions of mentioned in the Tender remain Unchanged.

[Signature]
20/5/25

Sr. Procurement-cum Store Officer (I/c)
AIIMS, Bhubaneswar